

PDF hosted at the Radboud Repository of the Radboud University Nijmegen

The following full text is a publisher's version.

For additional information about this publication click this link.

<http://hdl.handle.net/2066/81669>

Please be advised that this information was generated on 2019-04-24 and may be subject to change.

Results of the first international round robin for the quantification of urinary and plasma hepcidin assays: need for standardization

Joyce J.C. Kroot,¹ Erwin H.J.M. Kemna,¹ Sukhvinder S. Bansal,² Mark Busbridge,³ Natascia Campostrini,⁴ Domenico Girelli,⁴ Robert C. Hider,² Vasiliki Koliarakis,⁵ Avgi Mamelaki,⁵ Gordana Albina,⁶ Naohisa Tomosugi,⁷ Chris Tselepis,⁸ Douglas G. Ward,⁸ Tomas Ganz,^{6,9} Jan C.M. Hendriks,¹⁰ and Dorine W. Swinkels¹

¹Department of Clinical Chemistry, Radboud University Nijmegen Medical Centre, Nijmegen, The Netherlands; ²Pharmaceutical Sciences Division, King's College London, London, UK; ³Department of Clinical Chemistry, Imperial College HealthCare NHS Trust, Hammersmith Hospital Campus, London, UK; ⁴Department of Clinical and Experimental Medicine, Section of Internal Medicine, University of Verona, Verona, Italy; ⁵Laboratory of Molecular Biology & Immunobiotechnology, Department of Biochemistry, Hellenic Pasteur Institute, Athens, Greece; ⁶Intrinsic Life Sciences, La Jolla, CA, USA; ⁷Division of Advanced Medicine, Medical Research Institute/Division of Nephrology, Kanazawa Medical University, Ishikawa, Japan; ⁸CRUK Institute for Cancer Studies, University of Birmingham, Birmingham, UK; ⁹Department of Medicine, David Geffen School of Medicine at UCLA, LA, USA, and ¹⁰Department of Epidemiology and Biostatistics, Radboud University Nijmegen Medical Centre, Nijmegen, The Netherlands

ABSTRACT

The recently discovered iron regulatory peptide hormone hepcidin holds promise as a novel biomarker in iron metabolism disorders. To date, various mass spectrometry and immunochemical methods have been developed for its quantification in plasma and urine. Differences in methodology and analytical performance hinder the comparability of data. As a first step towards method harmonization, several hepcidin assays were compared. Worldwide eight laboratories participated in a urinary and plasma round robin in which hepcidin was analyzed. For both urine and plasma: (i) the absolute hepcidin concentrations differed widely between methods, (ii) the between-sample variation and the analytical variation of the methods are similar. Importantly, the analytical variation as percentage of the total variance is low for all methods, indicating their suitability to distinguish hepcidin levels of different samples. Spearman correlations between methods were generally

high. The round robin results inform the scientific and medical community on the status and agreement of the current hepcidin methods. Ongoing initiatives should facilitate standardization by exchanging calibrators and representative samples.

Key words: hepcidin, iron, quality control.

Citation: Kroot JJC, Kemna EHJM, Bansal SS, Busbridge M, Campostrini N, Girelli D, Hider RC, Koliarakis V, Mamelaki A, Albina G, Tomosugi N, Tselepis C, Ward DG, Ganz T, Hendriks JCM, and Swinkels DW. Results of the first international round robin for the quantification of urinary and plasma hepcidin assays: need for standardization. Haematologica 2009; 94:1748-1752. doi: 10.3324/haematol.2009.010322

©2009 Ferrata Storti Foundation. This is an open-access paper.

Introduction

Hepcidin plays a central role in iron metabolism, and could become a useful biomarker for the diagnosis and monitoring of iron disorders.^{1,2} Progress in human studies of hepcidin in normal physiology and various disease states has been hampered by the limited availability of hepcidin assays. Assays have been developed on mass spectrometry platforms including surface enhanced laser desorption/ionization time-of-flight mass spectrometry (SELDI-TOF MS), matrix assisted laser desorption/ionization time-of-flight mass spectrometry (MALDI-TOF MS) and liquid chromatography tandem-MS techniques (LC-MS/MS). Some methods use an internal standard, either hepcidin analogs or bioactive hepcidin-25 synthesized with stable isotopes.³⁻¹⁰ Recently, immunochemical (IC) assays for hepcidin-25 have also

been developed, which comprise of competitive radioimmunoassays (RIA)¹¹ and enzyme-linked immunosorbent assays (ELISA).^{12,13} Currently there is no reference method for hepcidin measurements. Therefore, to increase comparability of hepcidin data across clinical studies we evaluated the levels, between-sample variation and the analytical variation of hepcidin assays on a panel of urine and plasma samples in a so-called *Round Robin*.¹⁴

Design and Methods

Study design and participants

A prospective repeated measurement design with 12 replicates was used to assess concordance in urine and plasma hepcidin

Acknowledgments: we thank Siem Klaver, technician in the Department of Clinical Chemistry at Radboud University Nijmegen in The Netherlands, for his support in the collection and logistics of the study.

Manuscript received on April 18, 2009. Revised version arrived on May 28, 2009. Manuscript accepted on June 22, 2009.

Correspondence: Dorine W. Swinkels, Department of Clinical Chemistry, Radboud University Nijmegen Medical Centre, P.O. Box 9101 6500 HB, Nijmegen, The Netherlands. E-Mail: D.Swinkels@akc.umcn.nl

analysis. Six and eight laboratories participated in the urinary and plasma analyses, respectively, in six different countries. The study was coordinated by the Department of Clinical Chemistry of the Radboud University Nijmegen Medical Centre. All laboratories performed 12 replicates, that consisted of triplicate assays of each sample on four consecutive days. The only information provided about the samples were urinary creatinine levels.

Specimens

Eight urine samples with a wide range of hepcidin concentrations were collected from healthy subjects (samples 1-5) and patients (samples 6-8) (October 2007) with informed consent according to the declaration of Helsinki. Synthetic hepcidin-25 (Peptide Institute Inc., Osaka, Japan, net hepcidin peptide weight is precisely determined by amino acid analysis after acid hydrolysis) was added to a final concentration of 13.1 nmol/mmol creatinine to a urine sample that by SELDI-TOF MS⁷ was found to have a hepcidin concentration below the lower limit of detection of 0.5 nmol/L (sample 5). Seven plasma pools were composed from hospitalized patient sample remnants (March 2008), so as to cover a wide variation in hepcidin levels (samples 9-15). All samples were centrifuged for 10 min at 2600 g, and immediately stored in aliquots at -80°C. Two weeks after collection and storage, the samples were shipped on dry ice to all participants, and measured within four weeks of receipt, except for method VI (see below), that was carried out in July 2008. All samples underwent only one freeze-thaw cycle before analysis to minimize changes that may differentially affect the various methods, among which are the formation of hepcidin aggregates and breakdown products.

Hepcidin methods

Characteristics of the methods used for the plasma and urine hepcidin measurements of the present study are schematically presented in Table 1.

Method I is based on SELDI-TOF MS. Samples were directly applied to hydrophilic Normal Phase chips (NP20 ProteinChip; Bio-Rad Laboratories). Synthetic 25-hepcidin (Peptide Institute Inc.) was used for external mass calibration.

Method II is also based on SELDI-TOF MS. Hepcidin was first extracted from the sample using Macro-Prep® CM

Support beads (Bio-Rad Laboratories). The extract was applied to NP20 chips (Bio-Rad Laboratories). Synthetic hepcidin-24 peptide (custom made, Peptide Institute Inc.) was used as an internal standard.

For method III, urine samples were de-salted using C8 Clinprot beads (Bruker Daltonik) and analyzed by MALDI-TOF MS. Plasma samples were assayed by SELDI-TOF-MS using Cu²⁺ loaded IMAC chips. Synthetic human hepcidin-25 (Peptide Institute Inc.) was used as an external standard in both assays.

Method IV is based on LC-MS/MS. [¹⁵N,¹³C₂]Gly^{12,20} hepcidin (heavy hepcidin) was added to plasma or urine as internal standard. Magnetic nanoparticles (Bruker Daltonik) were used to extract hepcidin from the samples followed by LC-MS/MS analysis using selected reaction monitoring of the triple charged precursor fragmenting to the double charged product ion for both heavy hepcidin and hepcidin.

Method V is a competitive ELISA. 96-well plates were coated with in house-prepared anti-human hepcidin antibody and biotinylated hepcidin-25 as tracer. Custom synthesized hepcidin-25 was used as an external standard.

Method VI is a competitive RIA using a ¹²⁵I labeled synthetic hepcidin-25 (Bachem) with an in-house rabbit anti-hepcidin polyclonal antibody (against hepcidin-KLH conjugate) using a secondary antibody-PEG assisted separation. Synthetic hepcidin-25 (Bachem) was used as an external standard.

Method VII is based on LC-MS/MS. Isotopic human synthetic hepcidin-25 (Peptide Institute Inc.) was used as an internal standard.

Method VIII is a competitive ELISA: plates were coated with an in-house prepared hepcidin-25-His peptide and a polyclonal antibody against recombinant (in-house prepared) hepcidin-25-His was used to establish competition. Horseradish peroxidase labeled anti-rabbit antibody was used as secondary antibody. Hepcidin-25-His was used as an external standard.

Statistical methods

The study was designed to compare hepcidin levels as well as the repeatability of the methods used for serum and urine, respectively. With respect to the repeatability, the magnitude of variation that exists between samples and between measurements of the same sample relative to

Table 1. Characteristics of methods used for plasma and urine hepcidin measurements.

Method	Principle	Method	Hepcidin extraction	Standard	Urine	Plasma
I	Mass spectrometry	SELDI-TOF MS	Normal phase	None	X	X
II	Mass spectrometry	SELDI-TOF MS	Weak cation exchange	Internal	Synthetic hepcidin-24	X
III	Mass spectrometry	MALDI-TOF MS SELDI-TOF MS	Reversed phase Immobilized metal affinity Chromatography	External External	Synthetic hepcidin-25 Synthetic hepcidin-25	X X
IV	Mass spectrometry	LC-MS/MS	Weak cation exchange	Internal	[¹⁵ N, ¹³ C ₂]Gly-12,20-hepcidin	X
V	Immunochemical	Competitive ELISA	None	External	Synthetic hepcidin-25	X
VI	Immunochemical	Competitive RIA	None	External	Synthetic hepcidin-25	X
VII	Mass spectrometry	LC-MS/MS	None	Internal	Synthetic hepcidin-25	X
VIII	Immunochemical	Competitive ELISA	None	External	Recombinant hepcidin-25-His	X

X: participation in hepcidin round robin for urine and/or plasma.

Table 2. Mean (\pm SD) urine and plasma hepcidin levels presented by sample ID and method.

Body fluid	Sample	Method I*		Method II		Method III		Method IV		Method V		Method VI		Method VII		Method VIII	
		Mean	SD	Mean	SD	Mean	SD	Mean	SD	Mean	SD	Mean	SD	Mean	SD	Mean	SD
Urine	1	0.04	(0.01)	< LLOD	–	0.6	(0.4)	1.0	(0.6)	1.5	(0.3)	0.2	(0.01)9	–	–	–	–
	2	0.2	(0.1)	0.4	(0.1)	4.4	(4.0)	2.2	(0.7)	14.1	(0.5)	3.2	(0.2)9	–	–	–	–
	3	0.5	(0.1)	0.6	(0.1)	1.9	(0.6)	2.4	(0.7)	19.0	(4.2)	2.3	(0.4)9	–	–	–	–
	4	0.6	(0.2)	1.9	(0.4)	7.0	(0.7)	3.7	(1.1)	105.8	(176.0)	8.3	(1.0)9	–	–	–	–
	5	1.2	(0.4)	12.0	(3.0)	27.1	(6.7)	14.7	(1.9)	115.4	(14.9)	12.1	(0.9)9	–	–	–	–
	6	2.7	(0.5)	48.6	(6.3)	2.7	(0.6)	51.2	(14.7)	861.6	(308.2)	14.3	(1.6)9	–	–	–	–
	7	6.3	(1.2)	35.1	(6.5)	14.4	(3.3)	52.9	(11.5)	659.6	(176.4)	12.9	(1.2)9	–	–	–	–
	8	11.6	(2.7)	184.8	(36.0)11	49.2	(8.7)	153.0	(21.9)	1646.9	(597.1)	13.7	(0.9)9	–	–	–	–
Plasma	9	< LLOD	–	< LLOD	–	0.2	(0.6)	1.1	(0.4)	2.4	(3.8)	0.5	(0.1)	0.3	(0.1)	61.1	(6.6)
	10	4.6	(1.8)	4.3	(0.3)	14.3	(2.7)	7.1	(0.6)	39.7	(10.7)	9.4	(0.5)	4.1	(0.7)	32.7	(4.9)
	11	4.1	(1.0)	5.6	(1.1)	16.2	(2.6)	7.2	(0.5)	61.3	(12.9)	9.9	(0.8)	4.6	(0.7)	19.4	(2.7)
	12	7.5	(3.2)	8.4	(0.6)	23.4	(3.4)	14.1	(0.7)	128.9	(46.2)	21.2	(1.4)	9.3	(1.7)	23.1	(2.8)
	13	8.7	(3.6)	7.6	(2.4)	24.9	(3.9)	15.0	(1.0)	100.9	(53.5)	23.3	(0.9)	10.1	(1.8)	43.3	(5.4)
	14	18.3	(5.6)	23.8	(3.6)	52.6	(9.0)	34.7	(1.6)	259.5	(37.0)	29.6	(1.5)	29.2	(5.9)	58.0	(5.2)
	15	25.1	(6.0)	28.2	(2.6)	59.9	(12.2)	35.7	(1.5)	279.7	(44.2)	27.2	(2.0)	29.3	(4.0)	52.6	(6.7)

n, number of measurements if different from 12 (i.e., 3 measurements on four different days); urine results in nmol/mmol creatinine, plasma results in nmol/L; 1 nmol hepcidin-25=2.789 μ g hepcidin-25; #, Results reported in arbitrary Units, e.g. Mint/mmol creatinine for urine and in Mint/L for plasma; < LLOD, below lower level of detection, for method I: signal/noise < 3, for method II : < 50 pM and < 0.5 nM for urine and plasma, respectively. Mass spectrometry methods are highlighted in black.

the total variation are of interest. Accordingly, we partitioned the total variance of each hepcidin method into the following components: i) the between-sample variance and ii) the analytical variance. Our design allowed us to divide the latter into three subcomponents: the between-day variance, the between triplicate variance and the residual analytical variance, i.e. the part of the analytical variance that cannot be attributed to the other two.

A linear mixed model was used to estimate these variance components of each method separately. The dependent variable was hepcidin outcome, and the independent random variables were: sample (plasma: 7 levels, urine: 8 levels), day (4 levels) and repeated measurement (3 levels). We found that the estimated percentage variance due to the triplicate measurements was very small and considerably smaller than the variance due to the between-day variation. Consequently, this term was omitted from the final model and not presented separately. The SD (absolute error), the CV (relative error) and the percentage variance relative to the total variance of each random variable are presented and for each method separately.

Results and Discussion

Mean hepcidin levels for all (Table 2) samples differ considerably between all methods.

Notably, hepcidin values of method V are relatively high, but trends and variability are similar to all other methods (Table 2). It is also of note that for this Round Robin study, we only used native urine and serum samples, except for one, i.e. urine sample 5 to which synthetic hepcidin was added to a final concentration of 13.1 nmol/mmol creati-

nine. We found that the mean hepcidin outcome of some methods (i.e. methods II, IV and VI) for this spiked urine sample are closer to this concentration than others. However, irrespective of whether the value assigned of the spiked hepcidin-25 by Peptide International is correct, these results can not simply be extrapolated to native urine or serum samples, since it is presently unclear whether the various methods evaluated behave differently for spiked and native samples. In other words, the so called commutability of the spiked samples with native clinical samples for the various methods is unknown and its assessment may be part of future standardization efforts for hepcidin measurements.¹⁵

In general, differences in hepcidin levels between methods might be due to: (i) the use of different calibration solutions with level assignments based on different techniques; (ii) to hepcidin aggregation of either the standard solution or the sample; or (iii) hepcidin binding to α_2 macroglobuline or albumine¹⁶ or (iv) the existence of three hepcidin isoforms hepcidin-25, 22 and 20. These four points may differentially affect IC and MS measurements, and urine and serum quantifications. More specifically, it was recently found that around 90% of the circulating hepcidin is bound to α_2 macroglobulin in the blood. These observations not only raise the question whether we should measure total, bound or unbound hepcidin, but also what the methods evaluated actually measure. The search for an answer to these questions is a new challenge for which much can be learned from the measurements of steroid and thyroid hormones. Another cause for differences between IC and MS methods is that IC methods lack the selectivity to distinguish hepcidin-25 from hepcidins-20 and -22. However, the implications of including the latter two isoforms in the

Table 3. Sample means and variations by method for urine and plasma hepcidin levels.

Body fluid	Method	Mean ALL sample	SD	Between-sample			Analytical variation			SD	CV (%)	% of total variance
				CV (%)	% of total variance	SD	CV (%)	% of total variance	Residual analytical variation			
Urine	I	#2.9	#4.1	139.9	93.4	#<0.1	<0.1	<0.1	#1.1	37.3	6.6	
	II	44.4	66.1	148.8	95.9	4.9	11.0	0.5	12.8	28.9	3.6	
	III	13.4	16.9	125.9	93.7	1.8	13.3	1.1	4.0	29.8	5.2	
	IV	35.1	52.3	148.9	96.3	<0.1	<0.1	<0.1	10.2	29.1	3.7	
	V	427.1	587.8	137.6	85.4	102.8	24.1	2.6	220.6	51.6	12.0	
	VI	7.9	5.7	72.2	97.2	0.6	7.2	1.0	0.8	9.9	1.8	
Plasma	I	#11.4	#8.4	73.6	81.3	#1.6	14.4	3.1	#3.7	32.2	15.6	
	II	13.0	10.3	79.1	96.0	0.4	3.2	0.2	2.1	15.8	3.9	
	0	27.4	21.3	77.9	92.1	0.7	2.6	0.1	6.2	22.6	7.6	
	IV	16.4	13.6	83.2	99.5	0.2	1.2	<0.1	1.0	6.0	0.5	
	V	124.6	107.0	85.8	89.4	26.1	20.9	5.3	26.0	20.8	5.3	
	VI	17.3	10.8	62.3	98.8	0.1	0.6	<0.1	1.2	6.8	1.2	
	VII	12.4	11.9	96.3	94.4	<0.1	<0.1	<0.1	2.9	23.4	5.6	
	VIII	41.5	16.7	40.2	90.9	3.1	7.5	3.1	4.3	10.4	6.0	

Between-sample, segment due to variation between samples; analytical variation, segment due to repeated measurements; SD, standard deviation, i.e. absolute error; CV: coefficient of variation, i.e. relative error; % of total variance, % of variance-segment to the total variance; ‡, sample means and SDs in nmol/mmol creatinine and nmol/L for urine and plasma, respectively; unless otherwise stated; #, mean and SD in Mint/mmol creatinine for urine and in Mint/L for plasma. Mass spectrometry methods are highlighted in black.

assay on the total hepcidin values reported in the various iron disorders are not yet known.

Most methods are similar in both analytical variation and between-sample variation (Table 3). Of note is that the between-sample CV is lower for plasma than for urine hepcidin. This might indicate that the difference between urine and plasma is not due to the method, but more to biological mechanism, e.g. the hepcidin excretion pathways. MS methodologies II, IV and VII, exploiting an internal standard, show slightly lower contribution of the analytical variance to the total variance compared to the other MS methods I and III. These findings corroborate the assumption that the use of an internal standard decreases the analytical variation of the MS-techniques. IC-method V shows a high between-sample CV in combination with a relatively high analytical CV. Furthermore, IC-methods VI and VIII express the lowest between-sample CV compared to all other methods. However, both methods also display low analytical variation. This relatively low between-sample variation of both the latter IC-methods illustrates the difficulties in the generation of specific antibodies for hepcidin.

Importantly, the contribution of the analytical variation to the total variance is low for all methods (Table 3), which indicates the potential suitability of all methods to distinguish hepcidin levels of different samples. However, of note is that the higher the analytical variation of a method the lower the probability that populations with only small differences in hepcidin levels can be distinguished, e.g. hepcidin levels of healthy controls from that of patients with low-grade inflammation due to the presence of features of the metabolic syndrome.

Spearman correlations between the individual sample mean hepcidin values obtained by most methods were generally high (range 0.62-1.00), except for correlations

with method VIII that are somewhat disappointing (range 0.04-0.18). This should be interpreted with caution due to the small number of samples analyzed (n=7 and n=8 for plasma and urine, respectively).

In summary, hepcidin levels reported by the various methods vary considerably but analytical variance is generally low and similar for all methods. We recommend further harmonization of the various hepcidin assays by: (i) introducing an internal standard for all the MS-based methods used for clinical studies; (ii) reaching consensus on level assignment and level adjustment of the calibrators used in every procedure; (iii) production of a calibrator that mimics patient sera and (iv) regular testing of shared samples and/or calibrators that are commutable and have been value assigned for quality control.

Authorship and Disclosures

JJCK and EHJMK designed research, performed statistical analysis, contributed analytical tools, interpreted data and wrote the paper. SSB, MB, NC, DG, RCH, VK, AM, GO, NT, CT, DGW and TG contributed analytical tools and edited the paper. JCMH, designed and performed statistical analysis, interpreted data and co-wrote the paper. DWS designed research and statistical analysis, interpreted the data, wrote the paper and co-ordinated the study.

DWS and HT steer the www.hepcidinanalysis.com initiative that serves the scientific community with hepcidin analysis in body fluids.

VK and AM have patents applications to disclose (ELISA hepcidin); GO and TG are employees of Intrinsic Life Sciences and have ownership of stocks to disclose. The other authors have no conflict of interests to declare.

References

- Ganz T. Hepcidin and its role in regulating systemic iron metabolism. *Hematology. Am Soc Hematol Educ Program* 2006;507:29-35.
- Kemna EH, Tjalsma H, Willems HL, Swinkels DW. Hepcidin: from discovery to differential diagnosis. *Haematologica* 2008;93:90-7.
- Tomosugi N, Kawabata H, Wakatabe R, Higuchi M, Yamaya H, Umehara H, et al. Detection of serum hepcidin in renal failure and inflammation by using ProteinChip System. *Blood* 2006;108:1381-7.
- Murphy AT, Witcher DR, Luan P, Wroblewski VJ. Quantitation of hepcidin from human and mouse serum using liquid chromatography tandem mass spectrometry. *Blood* 2007;110:1048-54.
- Bozzini C, Camprostrini N, Trombini P, Nemeth E, Castagna A, Tenuti I, et al. Measurement of urinary hepcidin levels by SELDI-TOF-MS in HFE-hemochromatosis. *Blood Cells Mol Dis* 2008;40:347-52.
- Murao N, Ishigai M, Yasuno H, Shimonaka Y, Aso Y. Simple and sensitive quantification of bioactive peptides in biological matrices using liquid chromatography/selected reaction monitoring mass spectrometry coupled with trichloroacetic acid clean-up. *Rapid Commun Mass Spectrom* 2007;21:4033-8.
- Swinkels DW, Girelli D, Laarakkers C, Kroot J, Camprostrini N, Kemna EH, Tjalsma H. Advances in quantitative hepcidin measurements by time-of-flight mass spectrometry. *PLoS ONE* 2008;3:e2706.
- Ward DG, Roberts K, Stonelake P, Goon P, Zampronio CG, Martin A, et al. SELDI-TOF-MS determination of hepcidin in clinical samples using stable isotope labelled hepcidin as an internal standard. *Proteome Sci* 2008;6:28.
- Bansal SS, Halket JM, Bomford A, Simpson RJ, Vasavda N, Thein SL, Hider RC. Quantitation of hepcidin in human urine by liquid chromatography-mass spectrometry. *Anal Biochem* 2008;384:245-53.
- Kobold U, Dulffer T, Dangl M, Escherich A, Kubbies M, Roddiger R, Wright JA. Quantification of hepcidin-25 in human serum by isotope dilution micro-HPLC-tandem mass spectrometry. *Clin Chem* 2008;54:1584-6.
- Ashby DR, Gale DP, Busbridge M, Murphy KG, Duncan ND, Cairns TD, et al. Plasma hepcidin levels are elevated but responsive to erythropoietin therapy in renal disease. *Kidney Int* 2009;75:976-81.
- Ganz T, Olbina G, Girelli D, Nemeth E, Westerman M. Immunoassay for human serum hepcidin. *Blood* 2008;112:3922-3.
- Koliarakis V, Marinou M, Vassilakopoulos TP, Vavourakis E, Tsochatzis E, Pangalis GA, et al. A novel immunological assay for hepcidin quantification in human serum. *PLoS ONE* 2009; 4:e4581.
- Jacobs EM, Hendriks JC, van Tits BL, Evans PJ, Breuer W, Liu DY, et al. Results of an international round robin for the quantification of serum non-transferrin-bound iron: Need for defining standardization and a clinically relevant isoform. *Anal Biochem* 2005;341:241-50.
- Müller MM. Implementation of reference systems in laboratory medicine. *Clin Chem* 2000;46:1907-9.
- Peslova G, Petrak J, Kuzelova K, Hrdy I, Halada P, Kuchel PW, et al. Hepcidin, the hormone of iron metabolism, is bound specifically to α -2-macroglobulin in blood. *Blood* 2009;113:6225-36.